



Critical evaluation of various methods of estimating foetal weight by ultrasound

Shripad Hebbar

Department of Obstetrics and Gynecology, Department of OBG, Kasturba Medical College, Manipal

Summary: This prospective study was conducted in a peripheral hospital of Kasturba Medical College; Manipal to know which sonographic method of estimation of foetal weight reliably predicts the birth weight. 100 patients were scanned thoroughly within 48 hours of delivery. Seven different models of ultrasonic weight estimation were analysed critically. It was found that the method of Hadlock2, predicted the birth weight more accurately than others. The average deviation from the actual birth weight (226gms), the percentile values of absolute error of difference were least with this method and it predicted maximum number of cases within 10% of actual birth weight (85%). Significant differences were observed between the predicted and actual birth weight in all other methods ($p < 0.01$). It was concluded that this method is superior predictor of birth weight compared to other six and is a method of choice to estimate the birth weight in term pregnancies where the measurements of foetal head is inaccurate either because of engagement or moulding, as it incorporates only FL and AC measurements which is not affected by these changes.

Key words: prediction of birth weight, ultrasonography

Aims and Objectives:

In the last two decades, various models have been designed by different investigators to predict foetal weight using ultrasound. The desired outcome is achieved by measuring different foetal anthropometrical parameters. These investigators have found that an approximate estimation of foetal weight may be made by measuring Biparietal diameter (BPD), Head circumference (HC), Abdominal circumference (AC), and Femur length (FL).

Different models of ultrasound estimation of foetal weight have been proposed by Birnholz, Deter, Hadlock, Jordaan, Shepard and Warsof of which Hadlock and Shepard methods are most popular. However, these methods have not been evaluated in Indian context. A study was undertaken to analyse the accuracy of these seven methods to predict birth weight.

Materials and Methods:

This study was conducted at Dr. T.M.A. Pai Hospital, Udupi which is a peripheral unit of Kasturba Medical College, between January 2000 to June 2000. This hospital mainly caters low-risk obstetric population from low and mid socioeconomic groups in the surrounding villages. The high-risk patients like those who are having preterm labour, are usually referred to the main hospital and hence only term pregnancies ($n=100$) were selected for this study.

The patients studied were in between the ages of 21 and 36 years, 43 were primigravidae and remainder (57) were multigravidae and none had more than 4 pregnancies. Most of the patients could be recruited to study because either they got admitted for safe confinement because they hailed from a distant village or because they were posted for elective cesarean section.

Patients were scanned within 48 hours of delivery, using "TOSHIBA CAPASEE" (Toshiba Electronics, Japan) ultrasound scanner. A 3.5 Mhz abdominal transducer was used to obtain biparietal diameter (BPD), occipitofrontal diameter (OFD), anteroposterior and transverse abdominal diameters (AD1&AD2) and femur length (FL) in centimeters.

Correspondence:

Dr Shripad Hebbar, MD
Associate Professor of Obstetrics and Gynaecology
Kasturba Medical College Manipal
INDIA
Email: drshripadhebbar@yahoo.co.in

Head circumference (HC) was obtained using the equation:

$$HC = \pi(BPD+OFD)/2$$

Similarly abdominal circumference (AC) was calculated using the formula:

$$AC = \pi(AD1+AD2)/2$$

Average abdominal diameter was calculated by using the equation:

$$AD = (AD1+AD2)/2$$

One of the methods (Birnholz) needed corrected BPD value which was obtained by

$$\text{Corrected BPD} = (BPD * OFD / 1.264)^{0.5}$$

Table 1 shows formulae in different methods to arrive at estimated weight by ultrasound:

Models	Equations
Birnholz³	$BW = (3.42928 * BPD * AD^2 + 41.218) / 1000$
Deter⁴	$\text{Log}_{10}(BW) = -2.014 + 0.211 * BPD + 0.057 * AC - 0.00403 * BPD * AC$
Hadlock¹	$\text{Log}_{10}(BW) = -1.5213 + 0.003343 * AC * FL + 0.001837 * BPD^2 + 0.0458 * AC + 0.158 * FL$
Hadlock²	$\text{Log}_{10}(BW) = -1.696 + 0.1938 * FL + 0.5281 * AC - 0.004 * FL * AC$
Jordaan⁵	$\text{Log}_{10}(BW) = -1.1683 + 0.0377 * AC + 0.095 * BPD - 0.0015 * AC * BPD$
Shepard⁶	$\text{Log}_{10}(BW) = -1.7492 + 0.166 * BPD + 0.046 * AC - 0.002646 * AC * BPD$
Warsof⁷	$\text{Log}_{10}(BW) = -1.599 + 0.144 * BPD + 0.032 * AC - 0.000111 * BPD^2 * AC$

Within 15 minutes of delivery, neonates were weighed on an electronic scale (with accuracy of ± 5 gms) and the actual weight of the neonate was compared with ultrasonically estimated foetal weight. The difference between the two was recorded as error in grams.

Statistical methods:

The primary objective of this study was to determine how accurately each of the methods predicted the actual observed birth weight and to compare them with each other. The accuracy of prediction was defined as the absolute difference between the predicted and observed weight. The data analysis was performed with the help of a personal computer using SSPS version 7.5 (Statistical Package for Social Sciences). As the distribution of absolute error was distinctly nongaussian, nonparametric test for paired data was performed using Wilcoxon's signed ranks method for paired data.

Descriptive statistics are shown in Table 2.

Methods	Mean Birth Weight (gms)	St.Deviation (gms)
Birnholz ³	3064	440
Deter ⁴	2738	390
Hadlock ¹	2767	383
Hadlock ²	2834	387
Jordaan ⁵	2760	366
Shepard ⁶	2714	397
Warsof ⁷	2580	383
ActualWeight	2874	499

Table 2 shows that the mean birth weight in Hadlock2 method is comparable to that of mean of actual observed birth weight.

Percentile values for absolute error of difference in different methods (in grams) are shown in Table 3. It can be seen that the percentile values for error are least with Hadlock2 model.

Models	5 th	10 th	25 th	50 th	75 th	95 th
Birnholz³	32.3	61.5	120.8	237.8	438.6	558.4
Deter⁴	15.3	26.2	66.1	163.2	283.3	597.3
Hadlock¹	19.6	30.5	81.8	153.5	270.9	537.9
Hadlock²	7.3	28.9	62.8	144.1	259.9	526.8
Jordaan⁵	11.2	40.7	71.5	157.8	286.1	578.8
Shepard⁶	12.2	23.9	70.5	176.1	282.7	628.8
Warsof⁷	19.4	51.8	123.9	276.7	406.9	761.9

The overall variation from actual birth weight in each model is given in Table 4 (Absolute values are considered for analysis.) It can be seen that the variation is least in Hadlock2 model.

Methods	Deviation from actual birth weight
Birnholz³	± 325 gms
Deter⁴	± 285 gms
Hadlock¹	± 260 gms
Hadlock²	± 226 gms
Jordaan⁵	± 280 gms
Shepard⁶	± 295 gms
Warsof⁷	± 386 gms

From table 4 it is evident that the deviation is least with Hadlock2 method.

Accuracy of a method to predict birth weight within $\pm 5\%$ and $\pm 10\%$ of the actual birth weight are shown in Table 5.

Table 5

Methods	$\pm 5\%$	$\pm 10\%$
Birnholz³	30	55
Deter⁴	42	75
Hadlock¹	45	78
Hadlock²	53	85
Jordaan⁵	46	74
Shepard⁶	41	73
Warsof⁷	28	53

It can be seen that Hadlock2 method predicted the birth weight in 85% of cases within $\pm 10\%$ of actual birth weight.

Table 6 shows the results of Wilcoxon's signed ranks test.

Table 6

Models	'z' values	Two tailed significance test (p value)	
Birnholz³	-5.869 ^a	<0.01	S
Deter⁴	-4.700 ^b	<0.01	S
Hadlock¹	-3.878 ^b	<0.01	S
Hadlock²	-1.840 ^b	>0.01	NS
Jordaan⁵	-3.772 ^b	<0.01	S
Shepard⁶	-5.580 ^b	<0.01	S
Warsof⁷	-8.187 ^b	<0.01	S

a Based on negative ranks.
b Based on positive ranks.

It is evident from Table 6 that the difference between the predicted foetal weight and the actual observed birth weight is highly significant except that of Hadlock2 and all methods under estimate birth weight except that of Birnholz's.

Discussion:

From this study it can be concluded that Hadlock's method using FL&AC is more accurate in predicting the birth weight in term foetuses. This may be due to the fact that towards the term, the head has the tendency to fix or engage and these results in error in measuring head size. The moulding of head toward the term may also contribute to this phenomenon.

In the present study, all methods except that of Birnholz underestimated the foetal weight. Even though Hadlock

method using FL & AC underestimated the foetal weight; it was close to the actual observed birth weight, compared to other methods. It also could predict birth weight in maximum no. of cases within $\pm 10\%$ of actual birth weight.

Conclusion:

The improved weight estimates obtained using the method of Hadlock is based on the fact that FL is related linearly to crown heal length and according to Jordaan, crown heal length effects birth weight more significantly than the head size. And also it is easy to reproduce correct FL measurements, where as head measurements may be affected by variation of the shape of the head, moulding. Difficulty is also experienced in obtaining valid measurements when the head is deeply engaged in the pelvis.

References:

1. Hadlock FP, Harrist RB, Sharman RS et al. Estimation of foetal weight with the use of head and femur measurements. A prospective study. *Am J Obstet Gynecol* 1985; 151:333-337
2. Hadlock FP, Harrist RB, Carpenter RJ et al. Sonographic estimation of foetal weight; the use of femur length in addition to head and abdomen measurements. *Radiology* 1984;150:535-540
3. Birnholz JC. An algorithmic approach to accurate ultrasonic foetal weight estimation. *Invest Radiol* 1986; 21:571-576
4. Deter RL, Hadlock FP, Harrist RB et al. Evaluation of three methods for obtaining foetal weight estimates using dynamic image ultrasound. *J Clin Ultrasound* 9:421, 1981
5. Jordaan HVF. Estimation of foetal weight by ultrasound. *J of Clin Ultrasound* 1983; 11:59-66
6. Shepard MJ, Richard VA, Berkowitz RL et al. An evaluation of two equations for predicting foetal weight by ultrasound. *Am J Obstet Gynecol* 1982; 142:47-54
7. Warsof SL, Gohari P, Berkowitz RL et al. The estimation of foetal weight by computer assisted analysis. *Am J Obstet Gynecol* 1977; 1128:881-892